

Sections



1| Temperature Effects on K_d Measurements



2| Contract KinExA Research



3| Literature Spotlight - Regeneron



4| Ask the Inventor



5| Geek Corner



6| Tips & Tricks



Since 1995 KinExA has been the most accurate and sensitive biosensor available to determine affinity and kinetics.

1

Temperature Effects on K_d Measurements



As a general rule of thumb, the K_d becomes tighter and kinetic values will decrease as temperature decreases. We are aware of only a few cases where this rule is violated. For the most accurate data, binding measurements should be made under conditions as close to their intended application as possible. It is our opinion that, in pharmaceutical development, affinity and kinetic measurements should be made at 37°C since this is the temperature at which the final therapeutic will be delivered.

In keeping pace with the latest trends in physiologically relevant data measurements, we're pleased to announce a temperature controlled chamber designed to accommodate any KinExA instrument, the **TC1000**.

With the TC1000 you can easily incubate samples and run experiments at any temperature between 4°C and 40°C. Temperature is automatically recorded for each sample run and interior cameras allow viewing inside the chamber without perturbing the inside environment. The wide range of temperatures allows estimation of the thermodynamic parameters of binding (change in enthalpy ΔH , change in entropy ΔS , and change in heat capacity ΔC_p).




Additional features include:

- An integrated Windows 10 computer and monitor installed with KinExA Pro Software.
- Custom-made lab bench that is acid and chemical resistant and doubles as a convenient workstation.
- Level Sensors that monitor the buffer and waste levels.
- Replaceable air filters protect the evaporator and condenser coils from dirt and dust, maintaining high efficiency and reducing service needs.
- A tough steel interior and careful attention to design details ensure your unit will last and maintain a safe environment for the instrument and samples.
- Temperature uniformity allows for exact temperature control, providing more confidence in analysis results.
- Includes a one year warranty.


Sapidyne is accepting pre-orders for a 10% discount until January 31, 2021. The units will be available first quarter of 2021.

2

Contract KinExA Research



The unprecedented effects of COVID-19 are leaving many researchers tasked with catching up on a heavy workload and Sapidyne is here to help. Our contract research team is available to perform lead characterization (including cell based work) or third party verification for projects moving forward to an IND submission, FDA approval, or clinical trial. Services can be requested via our website, scientist.com, or scienceexchange.com.



Literature Spotlight

3

Capture Surface & SPR

In a recent study, scientists at Regeneron describe how SPR capture surfaces can profoundly impact the binding kinetics that are measured for molecular interactions. In the study, [6] different human monoclonal antibodies were measured using [8] different molecules on the capture surface. Binding constants measured using SPR were shown to be strongly influenced by the choice of capture molecule and consistently weaker than those measured in solution using KinExA. Depending on the immobilization molecule, SPR based K_d values were from 3 to 2826 times weaker than those measured using KinExA.

Kamat, Vishal, et al. **"The impact of different human IgG capture molecules on the kinetics analysis of antibody-antigen interaction."** ScienceDirect, Analytical Biochemistry, 10 Jan. 2020, www.sciencedirect.com/science/article/pii/S0003269719310917. DOI: 10.1016/j.ab.2020.113580

4

Ask the Inventor




Q: In brief, what are the differences between KinExA and SPR?

A: I think the most significant practical difference is that KinExA is suitable for measuring binding to intact cells. KinExA measures binding of a drug (or candidate) to intact cells in any liquid matrix (media, buffer, serum, etc). It has been demonstrated on both over expressing engineered cell lines and native cells expressing endogenous proteins. The measurement of direct binding to cell receptors eliminates concerns that a purified, solubilized, protein may not bind the same as a native protein, not to mention the time and expense of purifying and solubilizing. To my knowledge, no one has used a commercial SPR instrument with cells.

With that said, not all targets are membrane proteins and soluble measurements remain important. KinExA measures binding of the unmodified drug to the unmodified target in solution. Here again the solute can be anything (e.g. serum, cell lysate, buffer). SPR requires immobilization of one binding partner to a chip surface. The immobilization can cause conformational changes in the molecule leading to a change in K_d or the surface itself can cause matrix-mediated artifact errors in the measured K_d .

Finally, while there are published examples of single digit pM K_d 's measured with SPR, most scientists don't trust it for K_d 's tighter than 100 pM. Many do not trust it to accurately measure anything tighter than 1 nM. On the other hand, KinExA data with high affinity binding is prolific and trusted in scientific literature. The tightest binding affinity published is 12 fM! Please see Tech Note 234, KinExA vs SPR for an expanded version of this answer with references.



Geek Corner

5

Software Shortcuts

On the timing setup page in the Draw Source column you may enter or change your sample set using commas to specify non contiguous sample lines or a dash to enter contiguous sample lines. Time, Volume, and Rate of the sample set header row will be copied down to all lines of the sample set. Below are examples for the KinExA 3000, 3200, 4000, and Autosampler sample sets.

(Note: The Autosampler works the same as the 3000/3200 except the rack number must be specified. The first digit of a 3 digit number specifies the rack number, and the last two digits designate the tube number in the rack. For example, a tube in the 38th position on the 3rd rack would be 338.)

Concentrations may be entered in a similar fashion into the Concentration column of either the Timing Setup page or the Binding Signals page. You may not mix arithmetic operators (*,/,+,-) within a single dilution string. However you may enter a dilution string on any row of the concentration column. The ^ symbol reverses the order of the dilution string. The number to the left of the arithmetic operation is always assumed to be concentration. If no units are given then nM is applied.

Designating sample well locations for a 96 well microtiter plate can be sampled across the plate or down as seen in **Table 1**. If sampling down the plate, designate the rack number, incorporate a colon, and then the wells the samples will span (e.g. 3:A1-B4). The sample selection moves down, then right to left, starting at the top right corner. If sampling across the plate, designate samples similar to a normal rack. Make sure the dual microtiter plate rack is being used.



For a more comprehensive description and additional shortcuts, refer to Tech Note 209 Software Shortcuts.

KinExA 3000/3200	Ascending Order	Descending Order
Inserting lines 1-13	Type 1-13 then Enter	Type 13-1 then Enter
Specifying buffer	Type B,1-13 then Enter	Type B,13-1 then Enter
Repeating lines in between sample sets	Type 1-3,4,4,4,5-13 then Enter	Type 13-5,4,4,4,3-1 then Enter
4000/Autosampler	Ascending Order	Descending Order
Using Standards Rack 1-6	Type 1-6 then Enter	Type 6-1 then Enter
Using Rack 1 tubes 1-6	Type 101-106 then Enter	Type 106-101 then Enter
Spanning Multiple Racks	Type 3, 111, 215-219, 312-314 then Enter	Type 314-312, 219-215, 111, 3 then Enter
Concentrations	Ascending Order	Descending Order
Inserting 2 fold dilution starting at 400 pM and ending at 0	Type 400pM/2,0^ then Enter or 0,195.31fM*2 then Enter	Type 400pM/2,0 then Enter
Add NSB	Type 400pM/2,0^,NSB then Enter	Type 400pM/2,0,NSB then Enter
96 Well Microtiter Plate	Down	Across
Specifying 16 well positions	3: A1-B4	301-316

Table 1: Shortcuts for designating sample sets and concentration.

6 | Tips & Tricks



Aluminum Tape

One of the benefits of using a 4000 or Autosampler is the ability to queue up multiple experiments. Sometimes these experiments run over a long period of unattended operation (i.e. overnight or through the weekend). This autonomous operation time introduces a potential for evaporation if samples are left uncovered. To remedy these effects, we use aluminum adhesive tape to cover our samples.




The sampling arm on the instruments have no trouble piercing through the tape and it is a better alternative to X-Pierce film, labeling tape, or Parafilm. These alternatives are more costly, temperature sensitive, and cause residue to buildup on the sampling arm or cause z-axis crashes.

When applying tape, cover the entire sample set and make sure it has contact with the rack. This will ensure the samples stay in the rack once the sipper tip pulls away from each tube. The figure below is an example of how to cover samples.

When using large volumes with long run times samples can be prepared, split into two sets, and covered to avoid long term exposure during duplicate runs. The sample will then only be exposed when being drawn through the system. This method can also be advantageous for small volumes on a microtiter plate rack as well as 37°C measurements.

Tape can be purchased in a variety of different sizes. Below is a table with the sizes we use, what tubes and racks they cover, and a website to source the tape. We like JVCC AF20 Aluminum Foil Tape the best.



Tape Size	Surface Size	Item #	Website
1"	5 & 15 mL Conical	AF20/150	https://www.findtape.com/-JVCC-AF20-Aluminum-Foil-Tape/p267/?id=267&tid=21
2"	50 mL Conical	AF20/250	
4"	Microtiter Plate Rack	AF20/450	

25th Anniversary

Sāpidyne
INSTRUMENTS



est. **1995**

2020 has been a journey. Thank you for all of your hard work and dedication during this difficult year. We would have liked to have a party to celebrate 25 years however COVID had other plans. In lieu of a get together, we would like to send you an Idaho made gift so you can celebrate with us from your home or office. Just send an email with your shipping address to info@sapidyne.com.